



SIEVE SHAKER-GYRATORY-FOR 20CM DIA SIEVES WITH DIGITAL TIMER

Salient Features

- Carries upto 7 sieves of 200mm diameter in single adaptor.
- The shaker is driven by a 1/4 HP motor through a reduction gear
- The sieve table does not rotate, but is inclined from the vertical axis.
- The direction of inclination changes progressively in clockwise direction.
- If the stop-pin below the table is removed, the shaker can have a rotary motion.
- In addition to this gyratory motion of the table, there is an upward and downward movement ensuring that each square cm of the sieve is utilized.
- A pair of rods and a holder are supplied along with the main unit.
- The holder can be fixed on top of the upper most sieves, and thus the sieve set is firmly held.
- Suitable for operation on 230 volts, 50 cycles single phase A.C. Supply.
- Fitted with Digital time switch adjustable from 1-99 minutes.
- Supplied without sieves and Lid-Pan.



Click here for working video : <https://www.youtube.com/watch?v=USDLtLH6g>

Technical Specification

- Overall Dimensions in Inches (L x W x H) : 22 x 13 x 18
- No. of sieves accommodated : 6 sieves + 1 set of Lid-Pan
- Make of the Motor : GODREJ LAWKIM
- No. of tapping from bottom : 85 to 90
- Timer : Digital - Presettable
- Timer Range : 0 TO 99 Minutes

Corporate Office:

A-1301, BVR Ek, Opp. Hotel Inder Residency,
Near Gujarat College, Ellisbridge,
Ahmedabad-380 006.
Phone : (079) 6621 1234
Website : www.eieinstruments.com

Calibration Laboratory:

Survey No. 1098, Nr. Mahakali Temple,
Opp. Primary School, Village Kubadthal,
Tal-Daskroi, Ahmedabad-382 433.
Phone: (079)-3520 8300
E-mail : info@eieinstruments.com

GST TIN No. : 24AABCE4018L1ZU
I.T. PAN No. : AABCE 4018L
IEC No. : 0813026598 Dt. 04/03/2014
CIN No. : U29199GJ2004PTC045078
TAN No. : AHME00572F



EIE INSTRUMENTS PVT. LTD.

(An ISO 9001:2015 Certified Company)

- Power Supply : 220/230 Volts, 50 HZ, single Phase, AC
- Weight : 52 Kg Approx.

Corporate Office:

A-1301, BVR Ek, Opp. Hotel Inder Residency,
Near Gujarat College, Ellisbridge,
Ahmedabad-380 006.
Phone : (079) 6621 1234
Website : www.eieinstruments.com

Calibration Laboratory:

Survey No. 1098, Nr. Mahakali Temple,
Opp. Primary School, Village Kubadthal,
Tal-Daskroi, Ahmedabad-382 433.
Phone: (079)-3520 8300
E-mail : info@eieinstruments.com

GST TIN No. : 24AABCE4018L1ZU
I.T. PAN No. : AABCE 4018L
IEC No. : 0813026598 Dt. 04/03/2014
CIN No. : U29199GJ2004PTC045078
TAN No. : AHME00572F