




INITIAL SURFACE ABSORPTION TEST APPARATUS (ISAT)



AS PER BS 1881 PART 5:

This apparatus is used to assess the surface absorption characteristics of concrete. The rate of flow of water per unit area into a concrete surface when subjected to a constant head of 200 mm is measured. The unit consists of a capillary tube mounted on a scale, a water reservoir & connecting tubes. Easy to use, mounted on a stand. As per BS 1881: Part 208:1996

Weight : 30 kg Approx.

Technical Specification

- Overall Dimensions : 690 X 430 X 640(L X W X D) mm
- Weight (Approx.) : 30 kg

Corporate Office:

A-1301, BVR Ek, Opp. Hotel Inder Residency,
Near Gujarat College, Ellisbridge,
Ahmedabad-380 006.
Phone : (079) 6621 1234
Website : www.eieinstruments.com

Calibration Laboratory:

Survey No. 1098, Nr. Mahakali Temple,
Opp. Primary School, Village Kubadthal,
Tal-Daskroi, Ahmedabad-382 433.
Phone: (079)-3520 8300
E-mail : info@eieinstruments.com

GST TIN No. : 24AABCE4018L1ZU
I.T. PAN No. : AABCE 4018L
IEC No. : 0813026598 Dt. 04/03/2014
CIN No. : U29199GJ2004PTC045078
TAN No. : AHME00572F