

ANESTHESIA WORKSTATION MODEL *iJA 21* DURAPLUS

TECHNICAL SPECIFICATIONS:

VENTILATION FEATURES

Volume Control -
Tidal Volume - 50 to 1500
Frequency - 6 to 60
PIP - 60 CM OF WATER
INSP. TIME - .5 - 4 SEC

PRESSURE
CONTROL -
PRESSURE LIMIT - 10 - 60 CM OF WATER

INSP TIME - .5 - 4 SEC

FREQUENCY - 6 TO 60

STANDBY FOR MANUAL

PEEP - 0 - 30 CM OF WATER

DISPLAY-
ACTUAL PRESSURE LIMIT
I:E RATIO

ALARM -
HIGH PRESSURE
LOW PRESSURE
SOURCE FAILURE

SPECIFICATIONS

5 TUBE ROTAMETER-
2 OXYGEN, 2 NITROUS, 1 AIR WITH HYPOXIA
GAURD.

OXYGEN FAILURE AUDIO VISUAL ALARM

NITROUS/AIR CUTOFF

SELECTEC BACKBAR
2 VAPORIZER ATTACHMENT

RUST PROOF FIBER BODY

2 BIG DRAWER

WHEEL- 2x2
EASY MOVEMENT

CYLINDER ATTACHMENT- 2
OXYGEN, 1 NITROUS

PIPELINE ATTACHMENT- OXYGEN,
NITROUS AND AIR



An ISO & CE Certified Company

A Product of Inventure Sales, Regd. Office : H-206, Naraina Vihar, New Delhi-110028
+91 9818558570 sales@inventuresales.com; website : invsales.in