



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Name NTMT METROLOGY, # 4, SIPCOT SHOPPING COMPLEX, HOSUR, KRISHNAGIRI, TAMIL NADU, INDIA

Accreditation Standard ISO/IEC 17025:2005

Certificate Number CC-2970 Page No. : 1 / 5

Validity 14/03/2019 to 13/03/2021 Last Amended on -

'In view of the transition deadline for ISO/IEC 17025:2017, the validity of this accreditation certificate will cease on 30.11.2020.'

S.No	Discipline / Group	Quantity Measured/ Instrument	Range / Frequency	* Calibration Measurement Capability(±)	Remarks
Permanent Facility					
1	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Bevel protector, Combination Set L.C.: 5 min	0deg to 90deg to 0deg	3.5 arc min	Using Sine bar, Slip gauge by Comparison Method
2	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Bore gauge (transmission only) L.C.: 1 µm Dial range : Ø6-600 mm Probing Range : Up to 1mm	0 to 1 mm	5.8 µm	Using Electronic Dial Calibration Tester by Comparison Method based on JIS B 7515
3	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Calipers (Vernier / Dial / Digital) L.C.: 0.01 mm	0 to 1000 mm	13.2 µm	Using Caliper Checker / Grade "0" Gauge Blocks by Comparison Method based on IS 3651 (Part 1 & 2)
4	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Coating Thickness foils	10µm to 694 µm	2.8 µm	Using Digital Micrometer by Comparison Method
5	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Comparator Stand Flatness	400 mm X 400 mm	5.1 µm	Using Slip gauge, Lever dial gauge by Comparison Method



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Name NTMT METROLOGY, # 4, SIPCOT SHOPPING COMPLEX, HOSUR, KRISHNAGIRI, TAMIL NADU, INDIA

Accreditation Standard ISO/IEC 17025:2005

Certificate Number CC-2970 Page No. : 2 / 5

Validity 14/03/2019 to 13/03/2021 Last Amended on -

'In view of the transition deadline for ISO/IEC 17025:2017, the validity of this accreditation certificate will cease on 30.11.2020.'

S.No	Discipline / Group	Quantity Measured/ Instrument	Range / Frequency	* Calibration Measurement Capability(±)	Remarks
6	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Depth Caliper (Vernier / Dial / Digital) L.C.: 0.01 mm	0 to 300 mm	8.9 µm	Using Grade "0" Gauge Blocks & Long Slip gauge by Comparison Method based on IS 4213
7	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Depth micrometer(analog/digital)L.C.: 0.01 mm	0 to 300 mm	8.5 µm	Using Grade "0" Gauge Blocks & Depth Checker by Comparison Method based on BS 6468
8	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Dial Thickness GaugeL.C.: 0.01 mm	0 to 25 mm	7.4 µm	Using Grade "0" Gauge Blocks by Comparison Method based on IS 2092
9	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Digital Indicator/LVDT/ Electronic Probe L.C.: 0.0001 mmL.C.: 0.001 mm	0 to 25 mm	9.4 µm	Using Electronic Dial Calibration Tester by Comparison Method based on IS 2092
10	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Engineer Square, Angle plate	0 to 300 mm	8.6 µm	Using Granite Square, Slip gauge by Comparison Method
11	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	External Micrometer(analog/digital) Lc 0.01 mm	0 to 1000 mm	13.4 µm	Using Grade "0" Gauge Blocks by Comparison Method based on IS 2967



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Name NTMT METROLOGY, # 4, SIPCOT SHOPPING COMPLEX, HOSUR, KRISHNAGIRI, TAMIL NADU, INDIA

Accreditation Standard ISO/IEC 17025:2005

Certificate Number CC-2970 Page No. : 3 / 5

Validity 14/03/2019 to 13/03/2021 Last Amended on -

'In view of the transition deadline for ISO/IEC 17025:2017, the validity of this accreditation certificate will cease on 30.11.2020.'

S.No	Discipline / Group	Quantity Measured/ Instrument	Range / Frequency	* Calibration Measurement Capability(\pm)	Remarks
12	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Feeler Gauge	0.03 mm to 1 mm	5.9 μ m	Using Digital Micrometer by Comparison Method based on IS 3179
13	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Internal Micrometer/Stick Micrometer/L.C.: 0.01 mm	0 to 200 mm	8.2 μ m	Using Grade "0" Gauge Blocks by Comparison Method based on IS 2966
14	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Lever type Dial gauge L.C.: 0.002 mm	0 to 0.2mm	10.8 μ m	Using Electronic Dial Calibration Tester by Comparison Method based on IS 11498
15	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Pistol Caliper L.C.: 0.1mm	0 to 60 mm	59.3 μ m	Using Grade "0" Gauge Blocks by Comparison Method IS 2092
16	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Plunger type Dial gauge (analog / digital) L.C.: 0.01 mm	0 to 25 mm	8.3 μ m	Using Electronic Dial Calibration Tester by Comparison Method based on IS 2092
17	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Snap Gauge	0 to 300 mm	7 μ m	Using Slip gauge Set Grade '0' by Comparison method as per IS:3455, IS:7859 & IS:3477



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Name NTMT METROLOGY, # 4, SIPCOT SHOPPING COMPLEX, HOSUR, KRISHNAGIRI, TAMIL NADU, INDIA

Accreditation Standard ISO/IEC 17025:2005

Certificate Number CC-2970 Page No. : 4 / 5

Validity 14/03/2019 to 13/03/2021 Last Amended on -

'In view of the transition deadline for ISO/IEC 17025:2017, the validity of this accreditation certificate will cease on 30.11.2020.'

S.No	Discipline / Group	Quantity Measured/ Instrument	Range / Frequency	* Calibration Measurement Capability(\pm)	Remarks
18	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Sprit Level/Electronic Level (base)L.C. :5 μ m / m	Up to 200 mm	8.2 μ m/m	Using Electronic Level by Comparison Method
19	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Surface Plate	3000*3000 mm	0.76*sqrt(L+W)/100 μ m L & W in mm	Using Electronic Level by indirect Method
20	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	V BlockFlatness, Parallelism, Symmetry	0 mm to 300 mm	2.8 μ m	Using Mandrel, Slip gauge, Lever dial gauge by Comparison Method



National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



SCOPE OF ACCREDITATION

Laboratory Name NTMT METROLOGY, # 4, SIPCOT SHOPPING COMPLEX, HOSUR, KRISHNAGIRI, TAMIL NADU, INDIA

Accreditation Standard ISO/IEC 17025:2005

Certificate Number CC-2970 Page No. : 5 / 5

Validity 14/03/2019 to 13/03/2021 Last Amended on -

'In view of the transition deadline for ISO/IEC 17025:2017, the validity of this accreditation certificate will cease on 30.11.2020.'

S.No	Discipline / Group	Quantity Measured/ Instrument	Range / Frequency	* Calibration Measurement Capability(±)	Remarks
Site Facility					
1	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	1D/2D Electronic Height GaugeL.C.: 0.0001 µm	0 to 1000 mm	12.4 µm	Using Grade "0" Gauge Blocks & Long Steel Gauge Blocks by Comparison Method based on IS 2921
2	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Bench CentreParallelismCo-axiality	Up to 300 mm	8 µm	Using , Lever dial, Test Mandrel by Comparison Method
3	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Height gauge (Vernier / Dial / Digital)L.C.: 0.01 mm	0 to 1000 mm	13 µm	Using Grade "0" Gauge Blocks & Long Steel Gauge Blocks by Comparison Method based on IS 2921
4	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Straight Edge Flatness	Up to 3000 mm	5.8 µm	Using Electronic level, Lever dial, Test Mandrel by Comparison Method
5	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Surface Plate	3000*3000 mm	0.76*sqrt(L+W)/100 µm L & W in mm	Using Electronic Level by indirect Method