

CES - 405

HIGH VOLTAGE BREAKDOWN TESTER 0 - 5KV/30mA AC

CROWN" High Voltage Breakdown Tester is useful for measurement of Breakdown Voltage of Electrical Transformer, Cables & Wires Insulation Material & Components Etc. The Output Voltage & Leakage Current Limit is Settable As Required & Can be Monitor on Digital Meters The Instrument is Also Available With 3 Digit Digital Timer Settable From 1-999 Sec.



SPECIFICATION

OUTPUT VOLTAGE (AC)

0 - 5KV

TEST CURRENT

30mA

CURRENT LIMIT	:	Settable from 10% to 100% through continuously Variable control
METER	:	Two Digital meters to monitor output voltage & Leakage Current i. Digital KV Meter 0 to 5.00KV AC Least Count : 0.01KV ii. Digital mA Meter 0 to 199.9mA Least Count : 0.1mA
ACCURACY	:	± 1% of Reading
INDICATIONS	:	Lamp Indication for 'HT ON' & OF'
OUTPUT CABLE	:	1.25(approx) Meter Long HV Cable with HV Probe fitted with Push Button for HT ON
POWER REQUIREMENT	:	230V AC
TIMER(OPTIONAL)	:	3 Di it, setable from 0 - 999sec

HV Tester Can be Made As Per Customers Specification

(SPECIFICATIONS & DESIGN SUBJECT TO CHANGE WITHOUT PRIOR NOTICE)