

CUBEI30t@uch

SED-RATE SYSTEM

> TECHNICAL CHARACTERISTICS

POWER SUPPLY 110 to 230 VAC (50 – 60Hz) MIXING OF SAMPLES in compliance with CLSI and ICSH requirements

ABSORBED ELECTRICAL POWER max 100 W

DIMENSIONS 450 x 390 x 297 mm (w x h x d)

DISPLAY 10,1" wide vertical type

WEIGHT 14 Kg

NOISE LEVEL < 80 Db

WORKING TEMPERATURE + 15° to + 35°C

BAR-CODE READER Automatic internal RELATIVE HUMIDITY from 20% to 80% without condensation

> INSTRUMENTS AND ACCESSORIES

TEST DEVICE NEXT 500 500 tests Ref. 10293 TEST DEVICE NEXT 1K 1.000 tests Ref. 10294 TEST DEVICE NEXT EX. 5.000 tests Ref. 10294	
TECT DEVICE NEVT EV	
TEST DEVICE NEXT 5K 5.000 tests Ref. 10296	
TEST DEVICE NEXT 10K 10.000 tests Ref. 10297	
ESR CONTROL CUBE 4 x 9 ml Ref. 10435	
ESR CONTROL CUBE 2 x 9 ml Ref. 10436	
EXTERNAL BAR-CODE READER 1 Ref. 20550510	0
THERMAL PAPER 1 Ref. 10403	

MIEGGE MIEGGE

DIESSE Diagnostica Senese SpA
Via del Pozzo, 5 - Loc. San Martino
53035 Monteriggioni (SI) Italy
Ph: +39 0577 319560/61 - Fax: +39 0577 318763 - salesoffice@diesse.it
www.diesse.it





SED-RATE SYSTEM



30 ESR RESULTS
IN 25 MIN WITH MODIFIED
WESTERGREN METHOD



www.rajbiosis.com

+91-998-3333-469

INNOVATIVE CLINICAL DIAGNOSTIC SYSTEMS



CUBE 130 to auch

SED-RATE SYSTEM



FEATURES

- BATCH/CONTINUOS WORKING MODE
- TOUCH SCREEN INTERFACE
- REDUCED SIZE
- NO SAMPLE CONSUMPTION
- NO WASTE FLUIDS
- SAFETY







TEST PERFORMED ON EDTA/CBC SAMPLES

The test is performed on blood samples collected in the same top lavender tubes used for full blood count.

USER FRIENDLY

- Up to 30 samples simultaneously
- · Internal bar-code reader
- Ease of use
- · Reduction of turn-around time
- Standardization
- Internal mixing of samples by complete inversion of the tubes
- 10 inches colour touch screen
- · Innovative graphic interface
- 3 working modes: batch, continuos, mixed
- · Fast Turn-Around Time (TAT)
- · 30 results in less than 25 minutes



ENVIRONMENTALLY FRIENDLY

No production of waste materials, without extra-cost for liquid waste disposal.



ANALYZE DIFFERENT AND NON-DEDICATED TUBES